


Gold Conjugate –Data Sheet

Conjugate:	Monoclonal anti Biotin:	40nm Gold
Product Code:	CC.000407MAB40OD40	
Lot Number:	20010151	

Description of test	Document reference	BBI Specification	Result	Pass/Fail
Optical Density at 520nm at manufacture:	QCP064	40.0 – 44.0	43.4	Pass
Mean Particle Diameter:	QCP007	37.0-43.0	40.9nm	Pass
Coefficient of Variation of Particle Size:	QCP007	≤ 8.0%	≤8.0%	Pass
Buffer:	Conjugate is suspended in 2mM sodium Tetraborate at pH 7.2 containing 0.095% Sodium Azide.			
Date of Manufacture:	21 JAN 2020			

Data sheet compiled by: 

Date: 22 JAN 2020

Data reviewed by: 

Date: 22 Jan 2020

Print Date: 22 Jan. 20

Page 1 of 1

USA: 1-207-797-5454 | International: +44 (0)1495 363000 | info@bbisolutions.com | www.bbisolutions.com