

Gold Conjugate –Data Sheet

Conjugate:	Monoclonal anti Biotin:40nm	40nm Gold
Product Code:	CC.000407MAB40OD40	
Lot Number:	21010017	

Description of test	Document reference	BBI Specification	Result	Pass/Fail
Optical Density at 520nm at manufacture:	QCP064	40.0 – 44.0	40.3	Pass
Mean Particle Diameter:	QCP007	37.0 - 43.0 nm	41.8 nm	Pass
Coefficient of Variation of Particle Size:	QCP007	≤ 8.0%	≤8.0%	Pass
Buffer:	2mM Borax pH 7.2			
Date of Manufacture:	05 JAN 2021			

Data sheet compiled by: Seun Seriki

Date: 06 Jan 2021

Data reviewed by: *n. wat*

Date: *06 JAN 2021*

Print Date: 06 Jan. 21

Page 1 of 1

USA: 1-207-797-5454 | International: +44 (0)1495 363000 | info@bbisolutions.com | www.bbisolutions.com