

CERTIFICATE OF ANALYSIS

DEPLETED SERUM/PLASMA

T3/T4 Depleted Processed Serum

Code No.	SF231-7
Lot No.	

RECEIVER INFORMATION

Expiry Date

Manufacture Date:

Storage Temperature Below -15°C.

Storage Notes Avoid repeat freeze thaw cycles.

Shipping Notes Dry ice recommended.

PRODUCT INFORMATION

Source Processed serum.

Presentation Single homogenous batch, T3/T4 depleted processed serum filtered through 0.2µm membrane.

HEALTH AND SAFETY

Application For Research and Manufacturing Only.

Infectious Disease Tests

Starting material donor tested using FDA approved tests for:	Result
Hepatitis B Surface Antigen	
HIV 1 & 2 Antibodies	
Hepatitis C Antibodies	

Precaution No test can guarantee the absence of an infectious agent. Please handle as potentially hazardous. Appropriate PPE should be worn when handled.

Material Safety For further information and technical details, please download a Safety Data Sheet at www.bbisolutions.com or contact your BBI Account Manager.

ANALYSIS

ANALYTE	SPECIFICATION	RESULTS
Total Thyroxine (T4) level as determined Roche Modular	≤20 nmol/L	nmol/L
Total Triiodothyronine (T3) as determined by Roche Modular	≤1.0 nmol/L	nmol/L
Free Thyroxine (T4) level as determined by Roche Modular	≤1.0 pmol/L	pmol/L
Free Triiodothyronine (T3) as determined by Roche Modular	≤5.0 pmol/L	pmol/L
Bioburden	<10 cfu/ml	Pass/Fail

Name:		Position:	
Signed:		Date:	